

Super-resolution Imaging with Metamaterials for Cardiovascular Disease

Waterman, Naomi; Styles, Iain B.; Thomas, Steven G.; Zhang, Shuang

DOI:

[10.1364/BODA.2015.JT3A.10](https://doi.org/10.1364/BODA.2015.JT3A.10)

Document Version

Peer reviewed version

Citation for published version (Harvard):

Waterman, N, Styles, IB, Thomas, SG & Zhang, S 2015, Super-resolution Imaging with Metamaterials for Cardiovascular Disease. in *Optics in the Life Sciences*. Optical Society of America, pp. JT3A.10, Bio-Optics: Design and Application 2015, Vancouver, Canada, 12/04/15. <https://doi.org/10.1364/BODA.2015.JT3A.10>

[Link to publication on Research at Birmingham portal](#)

General rights

Unless a licence is specified above, all rights (including copyright and moral rights) in this document are retained by the authors and/or the copyright holders. The express permission of the copyright holder must be obtained for any use of this material other than for purposes permitted by law.

- Users may freely distribute the URL that is used to identify this publication.
- Users may download and/or print one copy of the publication from the University of Birmingham research portal for the purpose of private study or non-commercial research.
- User may use extracts from the document in line with the concept of 'fair dealing' under the Copyright, Designs and Patents Act 1988 (?)
- Users may not further distribute the material nor use it for the purposes of commercial gain.

Where a licence is displayed above, please note the terms and conditions of the licence govern your use of this document.

When citing, please reference the published version.

Take down policy

While the University of Birmingham exercises care and attention in making items available there are rare occasions when an item has been uploaded in error or has been deemed to be commercially or otherwise sensitive.

If you believe that this is the case for this document, please contact UBIRA@lists.bham.ac.uk providing details and we will remove access to the work immediately and investigate.

Super-resolution Imaging with Metamaterials for Cardiovascular Disease

Naomi Waterman^a, Iain Styles^{a,b}, Steve Thomas^c, Shuang Zhang^d.

^aPSIBS Doctoral Training Centre, ^bSchool of Computer Science, ^cCentre for Cardiovascular Science, ^dSchool of Physics & Astronomy, University of Birmingham, Edgbaston, Birmingham, UK, B15 2TT.

Author e-mail address: naw883@bham.ac.uk

Abstract: We propose a novel TIRF-like microscope design with a metamaterial lens. The microscope's structured evanescent field with subwavelength patterns will breach the diffraction limit. With it we will study cell-matrix interactions, important in cardiovascular disease.

OCIS codes: (110.0180) Imaging Systems; (160.3918) Materials; (000.1430) General.

1. Introduction

Optical microscopy, particularly fluorescence microscopy, is still a primary method of imaging cells and their organelles [1]. Much progress has been made in developing different optical microscopy techniques, leading to techniques such as Confocal Microscopy and TIRF Microscopy, in order to meet the multivariate demands of biomedical imaging [2]. A primary aim has been improving the resolution of optical microscopy techniques, which is fundamentally limited as defined by the Abbe limit [3], in order to image and understand deep subcellular processes.

TIRF microscopy is an optical technique that uses an evanescent field, produced by a wavefront undergoing total internal reflection in the sample's substrate, to illuminate a thin section ($\sim 100\text{-}200\text{nm}$) [2], of a sample with an exponentially decaying excitation field. This confined excitation improves upon the image resolution of optical microscopy by reducing background fluorescence. It is the principal technique for imaging the cell membrane and the process of endocytosis [1]. A promising super-resolution biomedical imaging technique is Structured Illumination Microscopy (SIM). The sample is illuminated with spatially structured excitation light – a periodic pattern. A series of such images is taken and are then processed to extract higher resolution information and generate a reconstruction with the improved resolution. However image acquisition is relatively slow [2]. The combination of TIRF and SIM has been investigated [4], and this project aims to build upon its progress to produce a super-resolution surface imaging technique.

The proposed microscope will use novel electromagnetic materials – metamaterials. Metamaterials can be defined as “an arrangement of artificial structural elements, designed to achieve advantageous and unusual electromagnetic properties” [5]. Therefore instead of a traditional material whose properties derive from its constituent atoms, a metamaterial derives its properties from its constituent structural elements or units. Consequently, materials with a negative refractive index are achievable, allowing for the potential of sub-diffraction limit optical imaging [6]. The proposed microscope will use a metamaterial lens to produce a TIRF-like evanescent field which will have a structured pattern, akin to SIM. As the fringes produced by the metamaterial are created by the surface plasmons, not an interference patterns in SIM, they are not diffraction limited, and the resulting excitation field is spatially confined as in TIRF, and can be structured as in SIM, but with a much higher spatial frequency.

2. Microscope Principles

The overall design of the proposed microscope is very similar to that of a brightfield microscope, the light travels in a straight line through the optical system, as opposed to the configuration of a TIRF microscope where the light must be incident on the sample substrate (usually a glass slide or cover slip) at an angle to produce an evanescent field. This is shown in cross-section in Figure 1a. . Light incident on the metalens produces a structured evanescent field from the far surface. The structured evanescent field then illuminates a thin slice of the sample, exciting fluorophores which then emit light which can be imaged in the far-field. The image of the fluorescence emission can then be reconstructed computationally to account for the structured excitation.





Figure 1. Schematic of: a - the proposed super-resolution microscope; b - the chosen metamaterial unit cell design, two 'H' shapes at $\pm 45^\circ$ to the normal. The metalens is formed from an array ($100 \times 100 \mu\text{m}$) of these unit cells.

3. Methodology - Simulations

Initial simulation work has been carried out in both MATLAB (Mathworks, Natick, MA) and CST Microwave Studio (CST, Framingham, MA). Simulations of the interaction of EM waves with different metamaterial designs and the image formation process have been conducted in CST-MWS to assess different potential designs of metalens. The electric field data from CST-MWS was then exported to and processed in MATLAB to determine the quality of the structured field produced. The design suggested by the simulation results is shown in Figure 1b, on the basis of which investigations into the feasibility of construction of a prototype metalens have begun. The proposed design consists of a unit cell design (two 'H' shapes at $\pm 45^\circ$ to the normal) cut from a gold film (Fig. 1b) which is reflected and tiled until the desired area is filled. The design is required to be symmetrical with a π phase difference between the two elements in order to produce the required structured field. Moreover the 'H' shape results in stronger field enhancement than a simpler shape, but is still relatively simple to manufacture.



5. Results

Preliminary work has been conducted in order to build a prototype metalens based upon the chosen design, see fig. 1b. Initial tests to determine the feasibility of the design have been conducted using electron beam lithography with test samples of PMMA mounted on a silicon wafer. Initial attempts have shown great promise that the design will be achievable.




Figure 3. SEM images of: a –the unit cell array of ‘H’ shapes; b – zoomed in SEM image of the unit cell array of ‘H’ shapes.

It can be seen, see fig. 3, that superficially the shapes are fairly close to the intended design, although they are less sharp than desired. The effects on the evanescent field of the imperfections of the ‘H’ shapes are being characterised via CST-MWS simulations. Moreover, the final design once made in gold will be fully characterized. Indeed it can also be seen that the dimensions of the ‘H’ shapes are of the correct order of magnitude ($\sim 151 \times 163 \text{ nm}$, should be $150 \times 150 \text{ nm}$). However they are generally slightly larger than desired, and are especially large in the centre of the ‘H’ ($\sim 76 \text{ nm}$ across, should be 50 nm). This is caused by overexposure of the PMMA due to secondary electrons.

6. Conclusions

The chosen metalens design has been shown in simulation to produce a strong structured evanescent field. The same design, but with a smaller unit cell demonstrated even stronger field enhancement, but may not be feasible to construct. The prototype metalens is practically feasible and is being trialled in gold. The accuracy of the construction of the metalens will be tested by imaging it with electron microscopy, AFM and FT-IR spectroscopy. Its imaging potential will be then tested using an inverted brightfield microscope and a nano-structured sample of known shape and dimensions. The images will then be reconstructed and compared to the simulation data. The metalens design will then be optimised and eventually tested on biological samples, notably platelets and their surface proteins, to investigate their link to cardiovascular disease.

7. References

- [1] A.L. Mattheyses, S.M. Simon and J.Z. Rappoport, “Imaging with total internal reflection fluorescence microscopy for the cell biologist,” *Journal of cell science* **123**, 3621-3628 (2010).
- [2] T.S. Tkaczyk, *Field Guide to Microscopy*, (SPIE Press; Bellingham, Wash, 2010).
- [3] D.B. Schmolze, C. Standley, K.E. Fogarty and A.H. Fischer, “Advances in microscopy techniques,” *Archives of pathology & laboratory medicine* **135**, 255-263 (2011).
- [4] M.G.L. Gustafsson, “Surpassing the lateral resolution limit by a factor of two using structured illumination microscopy,” *Journal of microscopy* **198**, 82-87 (2000).
- [5] W. Cai and V.M. Shalaev, *Optical metamaterials. Vol. 10*, (Springer, New York, 2010).
- [6] D.R. Smith, J.B. Pendry and M.C.K. Wiltshire, “Metamaterials and negative refractive index”. *Science* **305**, 788-792 (2004).